

Device Description	MAGILL intubation forceps (single use, non-sterile)
Basic UDI-DI	42515859NKMFDN5
UMDNS no.	10-678
GMDN code	33089
Classification	Class I, Rule 5, per annex VIII of the Regulation (EU) 2017/745

With the present declaration of conformity, we hereby declare, that the above-mentioned medical device meets the applicable* safety and performance requirements, defined in annex I of the Regulation (EU) 2017/745 as amended by Regulation (EU) 2020/561.

The conformity assessment procedure has been performed self-dependent according to Art. 52, paragraph 7, and Art. 19 of the Regulation (EU) 2017/745, for class I medical devices. We confirm that we have established and maintain a technical documentation according to annex II and III of the Regulation (EU) 2017/745, to the extent* of and as per the applicable standards:

- Quality management system (ISO 13485:2016)
- General safety & performance
- Risk management (ISO 14971:2019)
- Clinical evaluation (MEDDEV 2.7/1 Rev.4, MDCG 2020-5, MDCG 2020-6)
- Production processes & materials (ISO 7153-1:2016, ASTM F899-12b)
- Packaging, labelling & instructions of use (EN 1041:2008, ISO 15223-1:2016)
- Biocompatibility (Cytotoxicity ISO 10993-5:2009)
- Cleanliness (Bioburden ISO 11737-1:2018 guidelines)

According to Annex IV No. 2 of the Regulation (EU) 2017/745, we also declare that we, as the manufacturer, are solely responsible for issuing this EU Declaration of Conformity.

This declaration is valid until a revised declaration of conformity is issued.



Starking e.K.
SRN **DE-MF-000012725**
Rudolf-Diesel-Str. 18
78576 Emmingen-Liptingen, Germany



For Starkling e.K.
Emmingen-Liptingen, 21.03.2024